



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: AUG-21
Order Number: 138
Lot Number: 304-402119721-1
Cylinder Number: JB0012661

Customer: Hausnet S.R.L.

Use Before: 08/02/2025


Component	Specification (+/- 5%)	Analytical Result (+/- 2%)
Propane Air	1.05% VOL Balance	1.04% VOL Balance

Cylinder Size: 3.70 Cu. Ft.
Contents: 105 Liter

Valve: 5/8"-18UNF
Pressure: 1000 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:


Glenn Velez