



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: September 24, 2018
Order Number: D0001-20180692
Lot Number: JBI-248-100-15

Customer: Dastec SRL Argentina

Use Before: 09/24/2022

<u>Component</u>	<u>Specification (+/- 5%)</u>	<u>Analytical Result (+/- 2%)</u>
Isobutylene Air	100 PPM Balance	102.50 PPM Balance

Cylinder Size: 2.0 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:

Afton Briggs
Afton Briggs