



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: August 2019
Lot Number: GBL-135A-2.5-1
Cylinder Serial Number: GX0256188

Customer: Hausnet S.R.L.
Use Before: 08/13/2023

Component	Specification (+/- 10%)	Analytical Result (+/- 2%)
Methane Air	2.5% vol. Balance	2.468% vol. Balance

Cylinder Size: 3.70 Cu. Ft.
Contents: 105 Liter


Valve: 5/8" -18UNF
Pressure: 1200 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

TECHNICAL RECOMMENDATIONS:

For best results, use the following procedure when using Reactive Gas Mixtures and always use the shortest length possible of Teflon Tubing between the Cylinder Regulator Hose Barb and the Gas Detection Instrument. Turn the regulator on before connection is made with the cylinder, allowing any trapped air to be purged from the regulator. Be sure to turn the regulator off as soon as the regulator is fully connected. Always ensure delivery tubing is compatible with the Gas. Do not store this cylinder with the regulator installed due to possible leakage or long-term reaction with internal components of the regulator. Follow instrument manufacturer's instruction manual.

Analyst:


WILLIAM G. MAXWELL