



Calgaz Ltd
Unit 2 V22, Rosevale Road
Parkhouse Industrial Estate West
Newcastle-under-Lyme
Staffordshire
ST5 7EF
United Kingdom
info@calgaz.com
Phone: +44 1782 566 897

Certificate Of Composition WO249897 - 42

Part Code: C901030
6812104

Customer: Draeger Safety AG & Co KGaA
Customer Order Number: TC7NB4301979986

Cylinder No: 955-326381
Cylinder Valve: C10
Gross Weight: 1.5
Nett Weight: 0.12

<u>Component</u>	<u>Requested Values</u>	<u>Certified Values</u>
Methane	2.5 %	2.5 %
Air	Balance	Balance

Pressure: 1000 psi

Volume: 1.6 ltr

Size: 10AL

Please note all units are in mol % and mixture accuracy +/- 2%
Relative Mixtures traceable to standards calibrated at the National
Physical Laboratory, Teddington, Middlesex, England The UK
National Physical Laboratory (NPL) Standards are internationally
recognised and directly equivalent to the USA N.I.S.T Standards.

Manufactured Date: 14/04/2020

Valid Until: 14/04/2025

UN 1956 Compressed gas, n.o.s.

(Oxygen, Nitrogen Mixture)

Certified By: Andy Whitmore

All Gas Mixtures that are quoted with a balance of Air contain 20.9% (+/-2%) Oxygen relative to the overall Gas Mixture

The gases contained in this cylinder are not breathing gases.

DO NOT INHALE