



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: APR-22
Order Number: GAS-2223
Lot Number: 304-402383180-1

Customer: HAUSNET S.R. L.

Use Before: 04/15/2026

Component	Specification (+/- 5%)	Analytical Result (+/- 2%)
Carbon Monoxide Nitrogen	100 PPM Balance	100.3 PPM Balance

Cylinder Size: 3.64 Cu. Ft.
Contents: 103 Liter

Valve: 5/8" -18UNF
Pressure: 1000 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:


Glenn Velez