



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: August 23, 2019
Order Number: 74
Lot Number: IBJ-35-2.5-2

Customer: Hausnet S.R.L.

Use Before: 08/16/2023

Component	Specification (+/- 5%)	Analytical Result (+/- 2%)
Carbon Dioxide Nitrogen	2.5% vol. Balance	2.47% vol. Balance

Cylinder Size: 2.0 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:

Afton Eakins
Afton Eakins