



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: AUG-21
Order Number: 138
Lot Number: 304-402181902-1
Cylinder Number: A01521

Customer: Hausnet S.R.L.

Use Before: 08/02/2023

<u>Component</u>	<u>Specification (+/- 10%)</u>	<u>Analytical Result (+/- 2%)</u>
Carbon Monoxide	100 PPM	100.3 PPM
Methane	2.5% vol.	2.48% vol.
Hydrogen Sulfide	25 PPM	25 PPM
Oxygen	18% vol.	17.89% vol.
Nitrogen	Balance	Balance

Cylinder Size: 2.05 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:

Alton Eakins
Alton Eakins