



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
Fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: September 14, 2021
Order Number: D 0001-20210701
Lot Number: 304-402203237-1

Customer: DASTEC SRL USA LLC

Use Before: 09/14/2023


<u>Component</u>	<u>Requested Concentration</u>	<u>Analytical Result (+/- 2%)</u>
Carbon Monoxide	100 PPM	101 PPM
Methane	2.5% vol. (50% LEL)	2.49% vol.
Hydrogen Sulfide	25 PPM	26 PPM
Oxygen	18% vol.	17.9% vol.
Nitrogen	BALANCE	BALANCE

Cylinder Size: 2.0 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

Product composition verified by direct comparison to calibration standards traceable to N.I.S.T. weights and/ or N.I.S.T. Gas Mixture reference materials.

Analyst:


Glen Velez