



GASCO AFFILIATES, LLC

320 Scarlet Blvd
Oldsmar, FL 34677
(800) 910-0051
fax (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: APR-22
Order Number: GAS-2223
Lot Number: 304-402415381-1
Cylinder Number: A26421

Customer: Hausnet S.R.L.

Use Before: 04/15/2024

Component	Requested Concentration	Analytical Result (+/- 2%)
Carbon Monoxide	100 PPM	99.9 PPM
Methane	2.5% vol.	2.53% vol.
Hydrogen Sulfide	25 PPM	26 PPM
Oxygen	18% vol.	17.94% vol.
Air	Balance	Balance

Cylinder Size: 2.05 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:

Alton Eakins
Alton Eakins