



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: APR-21
Order Number: 128
Lot Number: 304-402167728-1

Customer: Hausnet S.R.L.
Use Before: 04/12/2025


Component	Specification (+/- 5%)	Analytical Result (+/- 2%)
Hydrogen Air	2% VOL Balance	2.05% VOL Balance

Cylinder Size: 3.70 Cu. Ft.
Contents: 105 Liter

Valve: 5/8"-18UNF
Pressure: 1200 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:


Glenn Velez