## **CERTIFICATE OF ANALYSIS**



ISO 9001:2015 CERTIFIED ISO 17034:2016 ACCREDITED ISO/IEC 17025:2017 ACCREDITED 1202 E Sam Houston Parkway S, Pasadena, TX 77503 Phone:(800) 548 2268 Fax:(713) 928 9961

PO Number:

4503131037

Manufactured For: MSA Safety Sales, LLC 1000 Cranberry Woods Drive Cranberry Township PA 16066 United States Of America Work Order No: 37448500 Certification Date: 13-Jul-2022

**Customer Part No:** 

10153845

Cylinder Size:

58DALS

Cylinder Type: ISO 7866

**Cylinder Content:** 

58 L (2 CU.FT.) @ 70 F (21 C) & 500 PSIG (3445Kpag)

Cylinder Lot No:

269184; 6 of 20 cylinders

**Cylinder Serial Nos:** 

BC432789,BC434999,BC435002,BC435305,BC474780,BC475011

**Expiration Date:** 

Jul 2024

Component	Nominal	Accuracy	Analytical Method
HYDROGEN SULFIDE	20 PPM (vol)	+/-10%	Automated Electrochemical Sensor
NITROGEN	BALANCE		
The state of the s			
		Mary Change to the Angelogy Control for June 2000 Control Cont	Francis deliber on the second second of the large formation of the property of the large formation of the large fo

The mixture was manufactured or transfilled from a standard which has been gravimetrically blended with traceability through NIST to the International System of Units (SI) balance. Balances are calibrated by a certified third party with certified NIST weights and NIST test numbers. Report Number: BU70910-111122. The uncertainty is expressed as an expanded uncertainty U=kuc with uc determined by experiment and a coverage factor k=2. The certified value +/-U is presented with a level of confidence of approximately 95%.

Quality Assurance Manager