



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
Fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: August 5, 2021
Order Number: D 0001-20210610
Lot Number: 304-402169724-1

Customer: Dastec SRL USA LLC

Use Before: 08/05/2023


<u>Component</u>	<u>Requested Concentration</u>	<u>Analytical Result (+/- 2%)</u>
Hydrogen Sulfide	25 PPM	26 PPM
Air	Balance	Balance

Cylinder Size: 2.0 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

Product composition verified by direct comparison to calibration standards traceable to N.I.S.T. weights and/ or N.I.S.T. Gas Mixture reference materials.

Analyst:


Brandon Brauo