



GASCO AFFILIATES, LLC.

320 Scarlet Blvd.
Oldsmar, FL 34677
(800) 910-0051
fax: (866) 755-8920
www.gascogas.com

CERTIFICATE OF ANALYSIS

Date: AUG-19
Order Number: Diego E-Mail
Lot Number: IBJ-161-20.9-4
Cylinder Number: GX0251147

Customer: Hausnet S.R.L.

Use Before: 08/13/2023

<u>Component</u>	<u>Specification (+/- 5%)</u>	<u>Analytical Result (+/- 2%)</u>
Oxygen	20.9% vol.	20.9% vol.
Nitrogen	Balance	Balance

Cylinder Size: 3.70 Cu. Ft.
Contents: 105 Liter

Valve: 5/8"-18 UNF
Pressure: 1200 psig

The calibration gas prepared by Gasco is considered a certified standard. It is prepared by gravimetric, or partial pressure techniques. The calibration standard provided is certified against Gasco's G.M.I.S. (Gas Manufacturer's Intermediate Standard) which is either prepared by weights traceable to the National Institute of Standards and Technology (NIST) or by using NIST Standard Reference Materials where available.

Analyst:

Afton Eakins
Afton Eakins