



Certificate of Analysis

Product name:

Zero Oxygen Solution Set (2 components)

or one month after mixing

Product code:

HI7040L

Lot number:

50004/18

Best use before:

Date of analysis:

January 2023

2018/02/01

Certified value, % Oz: 0.0± 0.1 at 25 °C

Method of standardization:

This quality product is standardized by lodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-0 B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	2496	January 2023
HI7040-2	2494	January 2023

Reference number: 01B82

OA manager:

Adela Odorhean

Certificate of Analysis

Product name:

Zero Oxygen Solution Set (2 components)

Product code:

HI7040L

Lot number:

50004/18

Restruse before:

anuary 2023

or one month after mixing

Date of analysis:

2018/02/01

Certified value, % O₂: 0.0± 0.1 at 25 °C

Method of standardization:

This quality product is standardized by lodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-0 B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	2496	January 2023
HI7040-2	2494	January 2023

Hanna Instruments Inc. 584 Park East Drive Woonsocket, RI 02895

Reference number: 01B82

OA manager:

Adela Odorhean

www.hannainst.com





Certificate of Analysis

Product name:

Zero Oxygen Solution Set (2 components)

or one month after mixing

Product code:

HI7040L

Lot number:

S0004/18

Best use before:

January 2023

Date of analysis:

2018/02/01

Certified value, % Oz: 0.0 ± 0.1 at 25 °C

Method of standardization:

This quality product is standardized by lodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-0 B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Components to a transfer and the same to a same to			
Component code:	Lot number:	Expiration Date:	
HI7040-1	2496	January 2023	
HI7040-2	2494	January 2023	

Reference number: 01B82

QA manager:

Adela Odorhean



Certificate of Analysis

Product name:

Zero Oxygen Solution Set (2 components)

or one month after mixing

Product code:

HI7040I

Lot number:

50004/18

Best use before:

January 2023

Date of analysis: 2018/02/01 Certified value, % O2: 0.0 ± 0.1 at 25 °C

Method of standardization:

This quality product is standardized by lodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-0 B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	2496	January 2023
HI7040-2	2494	January 2023

Reference number: 01B82

QA manager: Adela Odorhean

