

Certificate of Analysis

Product name: Zero Oxygen Solution Set (2 components)
Product code: HI7040L
Lot number: S0111/21
Best use before: September 2026 or one month after mixing
Date of analysis: 2021 / 10 / 06
Certified value, % O₂: 0.0 ± 0.1 @ 25 °C

Method of standardization:

This quality product is standardized by Iodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-O B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	6987	09 / 2026
HI7040-2	7032	10 / 2026

Reference number: 06J12

QA manager: Eugenia Tulbure

Certificate of Analysis

Product name: Zero Oxygen Solution Set (2 components)
Product code: HI7040L
Lot number: S0111/21
Best use before: September 2026 or one month after mixing
Date of analysis: 2021 / 10 / 06
Certified value, % O₂: 0.0 ± 0.1 @ 25 °C

Method of standardization:

This quality product is standardized by Iodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-O B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	6987	09 / 2026
HI7040-2	7032	10 / 2026

Reference number: 06J12

QA manager: Eugenia Tulbure

Certificate of Analysis

Product name: Zero Oxygen Solution Set (2 components)
Product code: HI7040L
Lot number: S0111/21
Best use before: September 2026 or one month after mixing
Date of analysis: 2021 / 10 / 06
Certified value, % O₂: 0.0 ± 0.1 @ 25 °C

Method of standardization:

This quality product is standardized by Iodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-O B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	6987	09 / 2026
HI7040-2	7032	10 / 2026

Reference number: 06J12

QA manager: Eugenia Tulbure

Certificate of Analysis

Product name: Zero Oxygen Solution Set (2 components)
Product code: HI7040L
Lot number: S0111/21
Best use before: September 2026 or one month after mixing
Date of analysis: 2021 / 10 / 06
Certified value, % O₂: 0.0 ± 0.1 @ 25 °C

Method of standardization:

This quality product is standardized by Iodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-O B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	6987	09 / 2026
HI7040-2	7032	10 / 2026

Reference number: 06J12

QA manager: Eugenia Tulbure