

Certificate of Analysis List

For request number 1430578

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2961742 Standard, 10 NTU	0322	N/A	N/A	StablCal sup TS sup0 Verification

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Verification Standard, 10 NTU**COMMODITY NUMBER: **2961742**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0322****11/19/2020****11/22/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.20 NTU

The expiration date is Nov 2022

Certified by

Scott Als
Analytical Services Chemist