

Certificate of Analysis List

For request number 1692463

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2961742-LM Standard, 10 NTU	1306	N/A	N/A	StablCal sup TS sup0 Verification

Total Enclosures: 1

***Certificate of Analysis***

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Verification Standard, 10 NTU**COMMODITY NUMBER: **2961742-LM**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1306****1/11/2022****1/18/2022**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	9.94 NTU

The expiration date is Nov 2023

Certified by

Scott Als
Analytical Services Chemist