


CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

AMETEK BROOKFIELD VISCOSITY REFERENCE STANDARD		
Viscosity Standard: 100		Lot Number: 011422
Certification/Issue Date: 01/14/22		Expiry Date: August 01, 2023
Viscosity	Temperature	
(centipoise)	(°C)	(°F)
97.8	25.0	77.00
Model	TEST INSTRUMENT	Serial Number
LVDV3T		8744156
Model	TEMPERATURE SENSOR	Serial Number
1502A		B86724

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE



David J. Moonay, Ph.D -Quality Engineer - Rheologist