

Certificate of Analysis List

For request number 1692467

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660242-LM NTU	2144	N/A	N/A	StablCal sup TS sup0 Standard, 100

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 100 NTU**COMMODITY NUMBER: **2660242-LM** MANUFACTURE DATE:LOT NUMBER: **A2144****6/27/2022**

DATE OF ANALYSIS:

6/27/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	95 to 105 NTU	101.0 NTU

The expiration date is Jul 2024

Certified by

Scott Als
Analytical Services Chemist