


CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Viscosity Standard: 1000		Lot Number: 012122	
Certification/Issue Date: 01/21/22		Expiry Date: August 01, 2023	
Viscosity		Temperature	
(centipoise)		(°C)	(°F)
983		25.0	77.00
Model	TEST INSTRUMENT		Serial Number
LVDV3T			8744156
Model	TEMPERATURE SENSOR		Serial Number
1502A			B86724

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE



David J. Moonay, Ph.D -Quality Engineer - Rheologist