



AMETEK Brookfield
INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
11 Commerce Blvd. | Middleboro, MA 02346
P: 508 940 6200 | F: 508 940 6262
www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

AMETEK BROOKFIELD VISCOSITY REFERENCE STANDARD		
Viscosity Standard 100		Lot Number 211321
Certification/Issue Date 01/12/23		Expiry Date July 01, 2025
Viscosity	Temperature	
(centipoise)	(°C)	(°F)
98.7	25.0	77.00
Model	TEST INSTRUMENT	Serial Number
LVDV3T		8744156
Model	TEMPERATURE SENSOR	Serial Number
1502A		B86724

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist