

Certificate of Analysis List

For request number 1430579

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660142	0294	N/A	N/A	StablCal sup TS sup0 Standard, 20 NTU

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 20 NTU**COMMODITY NUMBER: **2660142**

MANUFACTURE DATE:

LOT NUMBER: **A0294****10/28/2020**

DATE OF ANALYSIS:

11/2/2020

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.3 NTU

The expiration date is Oct 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist