

Certificate of Analysis List

For request number 1692465

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660142-LM	2095	N/A	N/A	StablCal sup TS sup0 Standard, 20 NTU

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 20 NTU**COMMODITY NUMBER: **2660142-LM** MANUFACTURE DATE:LOT NUMBER: **A2095****4/14/2022**

DATE OF ANALYSIS:

4/17/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.2 NTU

The expiration date is Apr 2024

Certified by

Scott Als
Analytical Services Chemist