



AMETEK Brookfield
INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
11 Commerce Blvd. | Middleboro, MA 02346
P: 508.946.6200 | F: 508.946.6262
www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Viscosity Standard: 30000		Lot Number: 112321	
Certification/Issue Date: 11/23/21		Expiry Date: July 27, 2023	
Viscosity		Temperature	
(centipoise)	(°C)	(°F)	
28800	25.0	77.00	
Model	TEST INSTRUMENT		Serial Number
RVDV3T			86009885
Model	TEMPERATURE SENSOR		Serial Number
1529			A56933

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist