



AMETEK Brookfield
INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
11 Commerce Blvd. | Middleboro, MA 02346
P: 508.946.6200 | F: 508.946.6262
www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Viscosity Standard: 5000		Lot Number: 122221	
Certification/Issue Date: 12/22/21		Expiry Date: August 01, 2023	
Viscosity		Temperature	
(centipoise)		(°C)	(°F)
4979		25.0	77.00
Model	TEST INSTRUMENT		Serial Number
LVDV-II+			8508226
Model	TEMPERATURE SENSOR		Serial Number
1502A			B1B961

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist