



Certificate of Analysis

Product name:

BUFFER SOLUTION

pH 4.01 ± 0.01 @ 25 °C (77 °F)

Product code:

HI7004L

Lot number: Best use before: 7373 January 2027

Date of analysis: Certified value:

2022.02.14 4.01 pH @ 25 °C (77 °F)

Method of standardization:

This quality product is standardized using a pH meter calibrated by means of two standard solutions prepared from below NIST standard reference materials:

> SRM 185i and

> SRM 186-I-q / SRM 186-II-q mixture

All primary standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty ${\it U}$ with a coverage of 2 and represents the 95% level of confidence.

Reference number: 25A05

OA manager:

QC_HI7004x_rev.1

Eugenia Tulbure

Certificate of Analysis

Product name:

BUFFER SOLUTION

pH 4.01 ± 0.01 @ 25 °C (77 °F)

Product code: Lot number:

HI7004L 7373

January 2027 Best use before: Date of analysis: 2022.02.14

4.01 pH @ 25 °C (77 °F) Certified value:

Method of standardization:

This quality product is standardized using a pH meter calibrated by means of two standard solutions prepared from below NIST standard reference materials:

> SRM 185i and

> SRM 186-I-g / SRM 186-II-g mixture

All primary standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage of 2 and represents the 95% level of confidence.

Reference number: 25A05

QA manager:

Eugenia Tulbure



QC_HI7004x_rev.1

Hanna Instruments Inc. 584 Park East Drive Woonsocket, RI 02895 www.hannainst.com

Hanna Instruments Inc. 584 Park East Drive Woonsocket, RI 02895 www.hannainst.com





Certificate of Analysis

Product name:

BUFFER SOLUTION

pH 4.01 ± 0.01 @ 25 °C (77 °F)

Product code:

HI7004L

Lot number:

7373

Best use before:

January 2027 2022.02.14

Date of analysis: Certified value:

4.01 pH @ 25 °C (77 °F)

Method of standardization:

This quality product is standardized using a pH meter calibrated by means of two standard solutions prepared from below NIST standard reference materials:

- > SRM 185i and
- > SRM 186-I-g / SRM 186-II-g mixture

All primary standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty \boldsymbol{U} with a coverage of 2 and represents the 95% level of confidence.

Reference number: 25A05

QA manager:

Eugenia Tulbure 🏂



QC_HI7004x_rev.1

Certificate of Analysis

Product name:

BUFFER SOLUTION

pH 4.01 ± 0.01 @ 25 °C (77 °F)

Product code:

HI7004L

Lot number:

7373 January 2027

Best use before: Date of analysis:

2022.02.14

Certified value:

4.01 pH @ 25 °C (77 °F)

Method of standardization:

This quality product is standardized using a pH meter calibrated by means of two standard solutions prepared from below NIST standard reference materials:

- > SRM 185i and
- > SRM 186-I-g / SRM 186-II-g mixture

All primary standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage of 2 and represents the 95% level of confidence.

Reference number: 25A05

QA manager:

Eugenia Tulbure



QC_HI7004x_rev.1