

Certificate of Analysis List

For request number 1692470

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2660542-LM NTU	2180	N/A	N/A	StablCal sup TS sup0 Standard, 800

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 800 NTU**COMMODITY NUMBER: **2660542-LM** MANUFACTURE DATE:LOT NUMBER: **A2180****7/28/2022**

DATE OF ANALYSIS:

8/1/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	800.5 NTU

The expiration date is Jul 2024

Certified by

Scott Als
Analytical Services Chemist