

Certificate of Analysis

Product name: CALIBRATION CHECKING SET FOR HI701

Product code: HI701-11

Lot number: SC0931/20

Best use before: December 2023

Date of analysis: 2020-11-20

Standard cuvette specifications @ 25 °C:

Standard cuvette	Lot number	Standard value [ppm]	Passed
HI701-11A	5913	0.00	✓
HI701-11B	5914	1.01 ± 0.05	✓

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not open standard cuvettes.

Instructions for use:

- 1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.*
- 2. At the C.1 prompt place HI701-11 A standard cuvette into the holder of the Checker and press operations button.*
- 3. At the C.2 prompt place HI701-11 B standard cuvette into the holder and press the operations button.*

Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT701-11 _ SC0931/20

QA manager: Andreea Man

