

Certificate of Analysis

Product name: CERTIFIED STANDARD CUVETTES
 Free Chlorine
 Product code: HI701-11
 Lot number: SC0615/23
 Best used before: August 2026
 Issuing date: 2023.07.31

Standard Cuvettes Specifications @ 25°C:

Standard cuvette code	Lot number	Standard value [ppm]	Analysis Results
HI701-11A	7774	0.00	Passed
HI701-11B	7775	1.00 ± 0.05	Passed

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not open standard cuvettes.

Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
 2. At the C.1 prompt place HI701-11A standard cuvette into the holder of the Checker and press operations button.
 3. At the C.2 prompt place HI701-11B standard cuvette into the holder and press the operations button.
- Acceptable results must be between 0.95 [ppm] and 1.05 [ppm] @ 25°C.

QA manager: Eugenia Tulbure/Engineer
 [Name/Title of signatory]

Signature: 