



AMETEK Brookfield
 INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
 11 Commerce Blvd. | Middleboro, MA 02346
 P: 508.946.6200 | F: 508.946.6262
 www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Viscosity Standard: 1000		Lot Number: 112723	
Certification/Issue Date: 11/27/23		Expiry Date: July 03, 2025	
Viscosity (centipoise) 992	Temperature		
		(°C)	(°F)
		25.0	77.00
TEST INSTRUMENT		Serial Number	
Model			86017552
LVDVNX			
TEMPERATURE SENSOR		Serial Number	
Model			B1B961
1502A			

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE. THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay

David J. Moonay, Ph.D -Quality Engineer - Rheologist