



AMETEK Brookfield
 INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
 11 Commerce Blvd. | Middleboro, MA 02346
 P: 508.946.6200 | F: 508.946.6262
 www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

| AMETEK BROOKFIELD VISCOSITY REFERENCE STANDARD | | |
|--|---------------------------|-----------------------------------|
| Viscosity Standard: 30000 | | Lot Number: 011224 |
| Certification/Issue Date: 01/12/24 | | Expiry Date: July 16, 2025 |
| Viscosity (centipoise) | Temperature | |
| | (°C) | (°F) |
| 30020 | 25.0 | 77.00 |
| Model | TEST INSTRUMENT | Serial Number |
| RVDV3T | | 86009885 |
| Model | TEMPERATURE SENSOR | Serial Number |
| 1502A | | B69204 |

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist