



AMETEK Brookfield
 INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
 11 Commerce Blvd. | Middleboro, MA 02346
 P: 508.946.6200 | F: 508.946.6262
 www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Lot Number: 112923

Expiry Date: June 27, 2025

Viscosity Standard: 5000

Certification/Issue Date: 11/29/23

Temperature

Viscosity

(centipoise)

4950

(°C)

25.0

(°F)

77.00

TEST INSTRUMENT

Model

LVDVNX

Serial Number

86017552

TEMPERATURE SENSOR

Model

1502A

Serial Number

B1B961

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist